



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## Certificate of Accreditation

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:*

***Midwest Scale Company***  
3445 Lonergan Drive, Rockford, IL 61109

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

**ISO/IEC 17025:2017**

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

***Mass, Force, and Weighing Devices Calibration***  
*(As detailed in the supplement)*

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen  
President

*Initial Accreditation Date:*

August 27, 2002

*Issue Date:*

March 26, 2024

*Expiration Date:*

April 30, 2026

*Accreditation No.:*

59146

*Certificate No.:*

L24-236

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: [www.pjllabs.com](http://www.pjllabs.com)*



# Certificate of Accreditation: Supplement

## Midwest Scale Company

3445 Lonergan Drive, Rockford, IL 61109  
 Contact Name: Jim Olejarski Phone: 815-713-9113

Accreditation is granted to the facility to perform the following calibrations:

### Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Class II Scales <sup>FO</sup>	1 g to 100 g	$(9.0 \times 10^{-4} + 2.23 \times 10^{-4}Wt) \text{ g}$	Class F Test Weights	NIST Handbook 44
	101 g to 200 g	$(4.1 \times 10^{-3} + 2.18 \times 10^{-3}Wt) \text{ g}$		
	201 g to 500 g	$(1.13 \times 10^{-1} + 5.55 \times 10^{-5}Wt) \text{ g}$		
	501 g to 1 000 g	$(1.19 \times 10^{-1} + 4.48 \times 10^{-5}Wt) \text{ g}$		
	1 001 g to 10 000 g	$(5.24 \times 10^{-2} + 1.11 \times 10^{-4}Wt) \text{ g}$		
	10 001 g to 32 000 g	$(6.16 \times 10^{-1} + 1.2 \times 10^{-4}Wt) \text{ g}$		
	32 001 g to 64 000 g	$(2.63 \times 10^{-1} + 1.13 \times 10^{-4}Wt) \text{ g}$		
Class III Scales <sup>FO</sup>	0.01 lb to 5 lb	$(1.2 \times 10^{-3} + 2.73 \times 10^{-5}Wt) \text{ lb}$		
	6 lb to 10 lb	$(8 \times 10^{-4} + 8.35 \times 10^{-5}Wt) \text{ lb}$		
	11 lb to 20 lb	$(1.13 \times 10^{-2} + 2.53 \times 10^{-5}Wt) \text{ lb}$		
	21 lb to 50 lb	$(1.11 \times 10^{-2} + 3.79 \times 10^{-5}Wt) \text{ lb}$		
	51 lb to 100 lb	$(9.6 \times 10^{-3} + 6.82 \times 10^{-5}Wt) \text{ lb}$		
	101 lb to 200 lb	$(6.7 \times 10^{-1} + 9.59 \times 10^{-5}Wt) \text{ lb}$		
	201 lb to 500 lb	$(1.1 \times 10^{-2} + 3.79 \times 10^{-5}Wt) \text{ lb}$		
	501 lb to 1 000 lb	$(5.72 \times 10^{-1} + 1.71 \times 10^{-5}Wt) \text{ lb}$		
	1 001 lb to 5 000 lb	$(5.32 \times 10^{-1} + 5.69 \times 10^{-4}Wt) \text{ lb}$		
	5 001 lb to 10 000 lb	$(3.42 \times 10^{-1} + 9.49 \times 10^{-5}Wt) \text{ lb}$		
	10 001 lb to 20 000 lb	$(2.01 \times 10^{-1} + 1.09 \times 10^{-5}Wt) \text{ lb}$		
	20 001 lb to 40 000 lb	$(1.36 + 9.49 \times 10^{-5}Wt) \text{ lb}$		
Class IIIL Scales <sup>FO</sup>	40 001 lb to 50 000 lb	$(4.31 + 7.8 \times 10^{-5}Wt) \text{ lb}$		
	50 001 lb to 120 000 lb Res 10 lb	$(9.46 + 6.9 \times 10^{-5}Wt) \text{ lb}$		
	50 001 lb to 120 000 lb Res 20 lb	$(21.8 + 4.03 \times 10^{-5}Wt) \text{ lb}$		

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represent the smallest measurement uncertainties attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is expressed at a confidence level of 95 % using a coverage factor  $k$  (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.



## *Certificate of Accreditation: Supplement*

### **Midwest Scale Company**

3445 Lonergan Drive, Rockford, IL 61109  
Contact Name: Jim Olejarski Phone: 815-713-9113

*Accreditation is granted to the facility to perform the following calibrations:*

3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location.
4. The presence of a superscript O means that the laboratory performs calibration of the indicated parameter onsite at customer locations.
5. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.
6. The term Wt represents weight in pounds or grams (including SI multiple and submultiple units) appropriate to the uncertainty statement.

